

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Chemtech **Client:** Weston Solutions
Lab Code: CHEM **Cas No:** Q1539 **SAS No :** Q1539 **SDG No:** Q1539
Matrix Spike - EPA Sample No : PB167101BSD **Datafile:** FG015476.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	204	102	78-117

LCS/LCSD % Recovery RPD : 1.0