

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1539

Client: Weston Solutions

Analytical Method: **8081B** Datafile : PL094591.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167076BS	alpha-BHC	0.5	0.45	ug/L	91				54	138	
	beta-BHC	0.5	0.46	ug/L	93				56	136	
	delta-BHC	0.5	0.47	ug/L	95				52	142	
	gamma-BHC (Lindane)	0.5	0.46	ug/L	91				59	134	
	Heptachlor	0.5	0.47	ug/L	93				54	130	
	Aldrin	0.5	0.47	ug/L	93				45	134	
	Heptachlor epoxide	0.5	0.48	ug/L	96				61	133	
	Endosulfan I	0.5	0.48	ug/L	96				62	126	
	Dieldrin	0.5	0.48	ug/L	96				60	136	
	4,4'-DDE	0.5	0.48	ug/L	96				57	135	
	Endrin	0.5	0.48	ug/L	96				60	138	
	Endosulfan II	0.5	0.49	ug/L	97				52	135	
	4,4'-DDD	0.5	0.49	ug/L	98				56	143	
	Endosulfan sulfate	0.5	0.50	ug/L	100				62	133	
	4,4'-DDT	0.5	0.49	ug/L	98				51	143	
	Methoxychlor	0.5	0.49	ug/L	99				54	145	
	Endrin ketone	0.5	0.51	ug/L	102				58	134	
	Endrin aldehyde	0.5	0.48	ug/L	97				51	132	
	alpha-Chlordane	0.5	0.47	ug/L	95				60	129	
	gamma-Chlordane	0.5	0.47	ug/L	95				56	136	
	Mirex	0.5	0.49	ug/L	98				51	127	