

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1539

Client: Weston Solutions

Analytical Method: 8081B

Datafile : PL094592.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167076BSD	alpha-BHC	0.5	0.43	ug/L	87	4			54	138	20
	beta-BHC	0.5	0.45	ug/L	89	4			56	136	20
	delta-BHC	0.5	0.46	ug/L	92	3			52	142	20
	gamma-BHC (Lindane)	0.5	0.43	ug/L	87	4			59	134	20
	Heptachlor	0.5	0.45	ug/L	90	3			54	130	20
	Aldrin	0.5	0.44	ug/L	89	4			45	134	20
	Heptachlor epoxide	0.5	0.46	ug/L	93	3			61	133	20
	Endosulfan I	0.5	0.47	ug/L	94	2			62	126	20
	Dieldrin	0.5	0.47	ug/L	94	2			60	136	20
	4,4'-DDE	0.5	0.46	ug/L	93	3			57	135	20
	Endrin	0.5	0.47	ug/L	95	1			60	138	20
	Endosulfan II	0.5	0.48	ug/L	96	1			52	135	20
	4,4'-DDD	0.5	0.48	ug/L	96	2			56	143	20
	Endosulfan sulfate	0.5	0.49	ug/L	98	2			62	133	20
	4,4'-DDT	0.5	0.48	ug/L	97	1			51	143	20
	Methoxychlor	0.5	0.49	ug/L	97	2			54	145	20
	Endrin ketone	0.5	0.51	ug/L	102	0			58	134	20
	Endrin aldehyde	0.5	0.48	ug/L	95	2			51	132	20
	alpha-Chlordane	0.5	0.46	ug/L	92	3			60	129	20
	gamma-Chlordane	0.5	0.46	ug/L	92	3			56	136	20
	Mirex	0.5	0.49	ug/L	98	0			51	127	20