

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WEST04
Lab Code: CHEM Case No.: Q1539 SAS No.: Q1539 SDG No.: Q1539
Instrument ID: BNA_F Calibration Date(s): 03/10/2025 03/10/2025
Calibration Time(s): 11:01 15:20

LAB FILE ID:		RRF2.5 = BF141897.D			RRF005 = BF141898.D		RRF010 = BF141899.D	
		RRF020 = BF141900.D			RRF040 = BF141901.D		RRF060 = BF141903.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
2-Fluorophenol		1.267	1.243	1.298	1.148	1.146	1.198	5.8
Benzaldehyde			1.021	0.975	0.778	0.716	0.853	15.8
Phenol-d6		1.623	1.593	1.644	1.442	1.471	1.526	6.0
Phenol		1.708	1.675	1.747	1.536	1.532	1.605	6.3
bis(2-Chloroethyl) ether		1.287	1.246	1.281	1.144	1.176	1.205	5.4
2-Chlorophenol		1.421	1.354	1.434	1.259	1.284	1.329	5.6
2-Methylphenol		1.080	1.050	1.122	1.008	1.079	1.057	3.7
2,2-oxybis(1-Chloropropane)		1.591	1.525	1.534	1.363	1.472	1.455	7.1
Acetophenone		0.514	0.493	0.513	0.460	0.469	0.479	5.8
3+4-Methylphenols		1.449	1.394	1.470	1.294	1.334	1.354	6.5
n-Nitroso-di-n-propylamine	1.020	1.024	1.000	1.043	0.913	0.991	0.980	5.3
Nitrobenzene-d5		0.345	0.356	0.379	0.348	0.363	0.355	3.5
Hexachloroethane		0.545	0.551	0.576	0.517	0.560	0.544	3.8
Nitrobenzene		0.347	0.351	0.375	0.346	0.359	0.353	3.1
Isophorone		0.647	0.620	0.655	0.600	0.648	0.628	3.5
2-Nitrophenol		0.150	0.159	0.182	0.175	0.188	0.173	7.9
2,4-Dimethylphenol		0.247	0.234	0.251	0.231	0.244	0.239	3.5
bis(2-Chloroethoxy) methane		0.411	0.406	0.416	0.372	0.397	0.394	4.6
2,4-Dichlorophenol		0.292	0.295	0.309	0.286	0.303	0.296	2.5
Naphthalene		1.104	1.075	1.101	0.984	0.973	1.027	6.2
4-Chloroaniline		0.383	0.378	0.393	0.351	0.352	0.363	6.1
Hexachlorobutadiene		0.211	0.211	0.221	0.203	0.214	0.213	2.5
Caprolactam		0.092	0.089	0.094	0.084	0.090	0.090	4.0
4-Chloro-3-methylphenol		0.334	0.331	0.349	0.318	0.334	0.331	3.1
2-Methylnaphthalene		0.746	0.709	0.733	0.653	0.667	0.687	6.1
Hexachlorocyclopentadiene		0.204	0.221	0.249	0.250	0.252	0.238	8.0
2,4,6-Trichlorophenol		0.388	0.389	0.402	0.398	0.401	0.398	1.9
2-Fluorobiphenyl		1.430	1.379	1.379	1.254	1.246	1.315	5.9
2,4,5-Trichlorophenol		0.396	0.391	0.422	0.389	0.415	0.403	3.0
1,1-Biphenyl		1.636	1.573	1.591	1.454	1.431	1.520	5.3
2-Chloronaphthalene		1.197	1.141	1.181	1.091	1.090	1.133	3.8
2-Nitroaniline		0.296	0.314	0.346	0.322	0.332	0.328	6.1
Dimethylphthalate		1.481	1.414	1.469	1.329	1.345	1.401	4.5
Acenaphthylene		1.765	1.740	1.778	1.635	1.595	1.681	4.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WEST04
Lab Code: CHEM Case No.: Q1539 SAS No.: Q1539 SDG No.: Q1539
Instrument ID: BNA_F Calibration Date(s): 03/10/2025 03/10/2025
Calibration Time(s): 11:01 15:20

LAB FILE ID:		RRF2.5 = BF141897.D			RRF005 = BF141898.D		RRF010 = BF141899.D	
		RRF020 = BF141900.D			RRF040 = BF141901.D		RRF060 = BF141903.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
2,6-Dinitrotoluene		0.277	0.279	0.305	0.285	0.296	0.292	4.1
3-Nitroaniline		0.291	0.295	0.316	0.293	0.301	0.296	3.6
Acenaphthene		1.236	1.195	1.231	1.146	1.165	1.181	3.4
2,4-Dinitrophenol			0.085	0.124	0.142	0.161	0.139	22.0
4-Nitrophenol		0.192	0.216	0.234	0.230	0.230	0.223	6.6
Dibenzofuran		1.849	1.787	1.791	1.614	1.596	1.688	6.9
2,4-Dinitrotoluene		0.365	0.382	0.402	0.375	0.383	0.381	2.9
Diethylphthalate		1.475	1.471	1.497	1.336	1.338	1.397	5.8
4-Chlorophenyl-phenylether		0.730	0.688	0.702	0.643	0.656	0.679	4.4
Fluorene		1.443	1.361	1.381	1.229	1.234	1.302	7.0
4-Nitroaniline		0.284	0.288	0.304	0.288	0.286	0.288	2.7
4,6-Dinitro-2-methylphenol			0.083	0.111	0.121	0.133	0.119	16.6
n-Nitrosodiphenylamine		0.675	0.651	0.671	0.620	0.638	0.647	3.0
2,4,6-Tribromophenol		0.234	0.236	0.259	0.250	0.265	0.254	5.7
4-Bromophenyl-phenylether		0.240	0.249	0.255	0.244	0.245	0.251	3.6
Hexachlorobenzene		0.262	0.267	0.284	0.269	0.271	0.275	3.9
Atrazine		0.216	0.213	0.193	0.164		0.198	10.7
Pentachlorophenol		0.130	0.151	0.172	0.177	0.186	0.170	12.7
Phenanthrene		1.155	1.151	1.138	1.038	1.032	1.080	5.9
Anthracene		1.173	1.139	1.139	1.046	1.036	1.084	5.9
Carbazole		0.996	0.994	0.985	0.908	0.892	0.935	5.9
Di-n-butylphthalate		1.320	1.321	1.319	1.177	1.175	1.240	6.5
Fluoranthene		1.234	1.226	1.228	1.108	1.059	1.146	7.7
Pyrene		1.695	1.651	1.727	1.611	1.823	1.730	5.4
Terphenyl-d14		1.343	1.309	1.360	1.276	1.417	1.353	3.6
Butylbenzylphthalate		0.629	0.656	0.702	0.655	0.717	0.677	4.7
3,3-Dichlorobenzidine		0.370	0.389	0.380	0.372	0.394	0.379	2.8
Benzo (a) anthracene		1.370	1.304	1.334	1.273	1.317	1.312	2.4
Chrysene		1.143	1.194	1.260	1.136	1.213	1.190	3.5
Bis (2-ethylhexyl) phthalate		0.897	0.912	0.966	0.904	0.952	0.934	3.4
Di-n-octyl phthalate		1.145	1.219	1.338	1.265	1.390	1.299	7.1
Benzo (b) fluoranthene		1.420	1.310	1.485	1.210	1.390	1.371	6.6
Benzo (k) fluoranthene		1.162	1.268	1.197	1.205	1.210	1.173	6.2
Benzo (a) pyrene		1.040	1.042	1.123	1.036	1.083	1.073	3.2
Indeno (1,2,3-cd) pyrene		1.097	1.142	1.308	1.287	1.472	1.294	10.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WEST04
Lab Code: CHEM Case No.: Q1539 SAS No.: Q1539 SDG No.: Q1539
Instrument ID: BNA_F Calibration Date(s): 03/10/2025 03/10/2025
Calibration Time(s): 11:01 15:20

LAB FILE ID:		RRF2.5 = BF141897.D		RRF005 = BF141898.D		RRF010 = BF141899.D		
		RRF020 = BF141900.D		RRF040 = BF141901.D		RRF060 = BF141903.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF060	RRF	% RSD
Dibenzo (a,h) anthracene		0.910	0.953	1.073	1.068	1.195	1.067	9.9
Benzo (g,h,i) perylene		0.912	0.949	1.061	1.065	1.167	1.055	9.2
1,2,4,5-Tetrachlorobenzene		0.615	0.604	0.626	0.592	0.606	0.609	2.1
1,4-Dioxane		0.490	0.482	0.536	0.504	0.501	0.502	3.4
2,3,4,6-Tetrachlorophenol		0.355	0.365	0.390	0.359	0.366	0.367	3.0