

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
 Data File : BF141953.D
 Acq On : 13 Mar 2025 21:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 10:40:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:46:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.502	0.488	2.8	90	-0.01
3	Pyridine	1.232	1.205	2.2	92	-0.01
4	n-Nitrosodimethylamine	0.587	0.553	5.8	88	-0.01
5 S	2-Fluorophenol	1.198	1.136	5.2	92	0.00
6	Aniline	1.505	1.380	8.3	88	0.00
7 S	Phenol-d6	1.526	1.422	6.8	92	0.00
8	2-Chlorophenol	1.329	1.247	6.2	92	0.00
9	Benzaldehyde	0.853	0.811	4.9	97	0.00
10 C	Phenol	1.605	1.483	7.6	90	0.00
11	bis(2-Chloroethyl)ether	1.205	1.114	7.6	91	0.00
12	1,3-Dichlorobenzene	1.435	1.352	5.8	92	0.00
13 C	1,4-Dichlorobenzene	1.452	1.386	4.5	93	0.00
14	1,2-Dichlorobenzene	1.367	1.293	5.4	92	0.00
15	Benzyl Alcohol	1.233	1.158	6.1	91	0.00
16	2,2'-oxybis(1-Chloropropane	1.455	1.252	14.0	86	0.00
17	2-Methylphenol	1.057	0.989	6.4	91	0.00
18	Hexachloroethane	0.544	0.512	5.9	92	0.00
19 P	n-Nitroso-di-n-propylamine	0.980	0.857	12.6	87	0.00
20	3+4-Methylphenols	1.354	1.231	9.1	89	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.479	0.447	6.7	89	0.00
23 S	Nitrobenzene-d5	0.355	0.345	2.8	90	0.00
24	Nitrobenzene	0.353	0.342	3.1	90	0.00
25	Isophorone	0.628	0.577	8.1	88	0.00
26 C	2-Nitrophenol	0.173	0.178	-2.9	93	0.00
27	2,4-Dimethylphenol	0.239	0.225	5.9	89	0.00
28	bis(2-Chloroethoxy)methane	0.394	0.364	7.6	89	0.00
29 C	2,4-Dichlorophenol	0.296	0.286	3.4	91	0.00
30	1,2,4-Trichlorobenzene	0.327	0.314	4.0	93	0.00
31	Naphthalene	1.027	0.970	5.6	90	0.00
32	Benzoic acid	0.210	0.217	-3.3	95	0.00
33	4-Chloroaniline	0.363	0.330	9.1	86	0.00
34 C	Hexachlorobutadiene	0.213	0.209	1.9	94	0.00
35	Caprolactam	0.090	0.081	10.0	88	0.00
36 C	4-Chloro-3-methylphenol	0.331	0.304	8.2	87	0.00
37	2-Methylnaphthalene	0.687	0.645	6.1	90	0.00
38	1-Methylnaphthalene	0.663	0.622	6.2	91	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	0.609	0.605	0.7	91	0.00
41 P	Hexachlorocyclopentadiene	0.238	0.213	10.5	76	0.00
42 S	2,4,6-Tribromophenol	0.254	0.243	4.3	86	0.00
43 C	2,4,6-Trichlorophenol	0.398	0.383	3.8	86	0.00
44	2,4,5-Trichlorophenol	0.403	0.398	1.2	91	0.00
45 S	2-Fluorobiphenyl	1.315	1.275	3.0	90	0.00
46	1,1'-Biphenyl	1.520	1.467	3.5	90	0.00
47	2-Chloronaphthalene	1.133	1.089	3.9	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
 Data File : BF141953.D
 Acq On : 13 Mar 2025 21:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 14 10:40:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 15:46:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.328	0.311	5.2	86	0.00
49	Acenaphthylene	1.681	1.594	5.2	87	0.00
50	Dimethylphthalate	1.401	1.310	6.5	88	0.00
51	2,6-Dinitrotoluene	0.292	0.282	3.4	88	0.00
52 C	Acenaphthene	1.181	1.117	5.4	87	0.00
53	3-Nitroaniline	0.296	0.273	7.8	83	0.00
54 P	2,4-Dinitrophenol	0.139	0.124	10.8	78	0.00
55	Dibenzofuran	1.688	1.577	6.6	87	0.00
56 P	4-Nitrophenol	0.223	0.208	6.7	80	0.00
57	2,4-Dinitrotoluene	0.381	0.364	4.5	86	0.00
58	Fluorene	1.302	1.205	7.5	87	0.00
59	2,3,4,6-Tetrachlorophenol	0.367	0.351	4.4	87	0.00
60	Diethylphthalate	1.397	1.298	7.1	86	0.00
61	4-Chlorophenyl-phenylether	0.679	0.639	5.9	88	0.00
62	4-Nitroaniline	0.288	0.265	8.0	82	0.00
63	Azobenzene	1.298	1.180	9.1	84	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.115	3.4	80	0.00
66 c	n-Nitrosodiphenylamine	0.647	0.639	1.2	86	0.00
67	4-Bromophenyl-phenylether	0.251	0.253	-0.8	87	0.00
68	Hexachlorobenzene	0.275	0.274	0.4	85	0.00
69	Atrazine	0.198	0.161	18.7	83	0.00
70 C	Pentachlorophenol	0.170	0.168	1.2	79	0.00
71	Phenanthrene	1.080	1.024	5.2	83	0.00
72	Anthracene	1.084	1.029	5.1	83	0.00
73	Carbazole	0.935	0.858	8.2	79	0.00
74	Di-n-butylphthalate	1.240	1.189	4.1	85	0.00
75 C	Fluoranthene	1.146	0.973	15.1	74	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
77	Benzidine	0.279	0.467	-67.4#	96	0.00
78	Pyrene	1.730	1.587	8.3	73	0.00
79 S	Terphenyl-d14	1.353	1.269	6.2	73	0.00
80	Butylbenzylphthalate	0.677	0.638	5.8	72	-0.01
81	Benzo(a)anthracene	1.312	1.234	5.9	71	0.00
82	3,3'-Dichlorobenzidine	0.379	0.416	-9.8	82	0.00
83	Chrysene	1.190	1.160	2.5	75	0.00
84	Bis(2-ethylhexyl)phthalate	0.934	0.882	5.6	72	-0.01
85 c	Di-n-octyl phthalate	1.299	1.355	-4.3	79	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	1.294	1.148	11.3	83	-0.01
88	Benzo(b)fluoranthene	1.371	1.194	12.9	91	0.00
89	Benzo(k)fluoranthene	1.173	1.097	6.5	84	0.00
90 C	Benzo(a)pyrene	1.073	1.017	5.2	91	-0.01
91	Dibenzo(a,h)anthracene	1.067	0.969	9.2	84	-0.02
92	Benzo(g,h,i)perylene	1.055	0.897	15.0	78	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031325\
Data File : BF141953.D
Acq On : 13 Mar 2025 21:09
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 14 10:40:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 15:46:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0