

Report of Analysis

Client:	Weston Solutions		Date Collected:	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	
Client Sample ID:	PB167097BSD		SDG No.:	Q1539
Lab Sample ID:	PB167097BSD		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141942.D	1	03/12/25 08:45	03/13/25 13:47	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
100-52-7	Benzaldehyde	21.2		3.90	8.00	10.0	ug/L
108-95-2	Phenol	43.0		0.91	4.00	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	42.8		0.81	4.00	5.00	ug/L
95-57-8	2-Chlorophenol	44.2		0.58	4.00	5.00	ug/L
95-48-7	2-Methylphenol	44.5		1.10	4.00	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	40.4		1.30	4.00	5.00	ug/L
98-86-2	Acetophenone	47.2		0.74	4.00	5.00	ug/L
65794-96-9	3+4-Methylphenols	44.0		1.10	8.00	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	41.0		1.40	2.50	2.50	ug/L
67-72-1	Hexachloroethane	42.7		0.65	4.00	5.00	ug/L
98-95-3	Nitrobenzene	43.6		0.76	4.00	5.00	ug/L
78-59-1	Isophorone	45.2		0.75	4.00	5.00	ug/L
88-75-5	2-Nitrophenol	45.9		1.80	4.00	5.00	ug/L
105-67-9	2,4-Dimethylphenol	54.8		1.90	4.00	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	42.7		0.68	4.00	5.00	ug/L
120-83-2	2,4-Dichlorophenol	43.8		0.52	4.00	5.00	ug/L
91-20-3	Naphthalene	42.1		0.50	4.00	5.00	ug/L
106-47-8	4-Chloroaniline	18.2		0.84	4.00	5.00	ug/L
87-68-3	Hexachlorobutadiene	44.1		0.54	4.00	5.00	ug/L
105-60-2	Caprolactam	48.9		1.10	8.00	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	43.1		0.59	4.00	5.00	ug/L
91-57-6	2-Methylnaphthalene	41.5		0.56	4.00	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	160	E	3.60	8.00	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	45.1		0.51	4.00	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	45.7		0.62	4.00	5.00	ug/L
92-52-4	1,1-Biphenyl	48.0		0.53	4.00	5.00	ug/L
91-58-7	2-Chloronaphthalene	43.7		0.61	4.00	5.00	ug/L
88-74-4	2-Nitroaniline	45.2		1.30	4.00	5.00	ug/L
131-11-3	Dimethylphthalate	42.5		0.61	4.00	5.00	ug/L

Report of Analysis

Client:	Weston Solutions		Date Collected:	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	
Client Sample ID:	PB167097BSD		SDG No.:	Q1539
Lab Sample ID:	PB167097BSD		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141942.D	1	03/12/25 08:45	03/13/25 13:47	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
208-96-8	Acenaphthylene	46.1		0.75	4.00	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	45.0		0.92	4.00	5.00	ug/L
99-09-2	3-Nitroaniline	26.8		1.10	4.00	5.00	ug/L
83-32-9	Acenaphthene	49.6		0.55	4.00	5.00	ug/L
51-28-5	2,4-Dinitrophenol	95.0	E	6.00	8.00	10.0	ug/L
100-02-7	4-Nitrophenol	97.3	E	2.40	8.00	10.0	ug/L
132-64-9	Dibenzofuran	42.0		0.61	4.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	47.0		1.20	4.00	5.00	ug/L
84-66-2	Diethylphthalate	42.7		0.69	4.00	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	43.6		0.68	4.00	5.00	ug/L
86-73-7	Fluorene	43.5		0.63	4.00	5.00	ug/L
100-01-6	4-Nitroaniline	44.2		1.50	4.00	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	49.9		2.90	8.00	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	44.8		0.58	4.00	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	44.3		0.40	4.00	5.00	ug/L
118-74-1	Hexachlorobenzene	47.0		0.52	4.00	5.00	ug/L
1912-24-9	Atrazine	59.1		1.00	4.00	5.00	ug/L
87-86-5	Pentachlorophenol	95.6	E	1.60	8.00	10.0	ug/L
85-01-8	Phenanthrene	45.0		0.50	4.00	5.00	ug/L
120-12-7	Anthracene	45.8		0.61	4.00	5.00	ug/L
86-74-8	Carbazole	44.5		0.72	4.00	5.00	ug/L
84-74-2	Di-n-butylphthalate	43.1		1.20	4.00	5.00	ug/L
206-44-0	Fluoranthene	43.3		0.82	4.00	5.00	ug/L
129-00-0	Pyrene	44.7		0.50	4.00	5.00	ug/L
85-68-7	Butylbenzylphthalate	43.7		1.90	4.00	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	29.8		0.93	8.00	10.0	ug/L
56-55-3	Benzo(a)anthracene	46.3		0.45	4.00	5.00	ug/L
218-01-9	Chrysene	44.3		0.44	4.00	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	42.4		1.60	4.00	5.00	ug/L
117-84-0	Di-n-octyl phthalate	43.5		2.30	8.00	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	39.4		0.49	4.00	5.00	ug/L

Report of Analysis

Client:	Weston Solutions		Date Collected:	
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169		Date Received:	
Client Sample ID:	PB167097BSD		SDG No.:	Q1539
Lab Sample ID:	PB167097BSD		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF141942.D	1	03/12/25 08:45	03/13/25 13:47	PB167097

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	45.8		0.48	4.00	5.00	ug/L
50-32-8	Benzo(a)pyrene	47.3		0.55	4.00	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	48.6		0.59	4.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	48.6		0.67	4.00	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	45.5		0.69	4.00	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	47.7		0.52	4.00	5.00	ug/L
123-91-1	1,4-Dioxane	41.8		1.00	4.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	43.9		0.72	4.00	5.00	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	132		19 - 119		88%	SPK: 150
13127-88-3	Phenol-d6	129		10 - 130		86%	SPK: 150
4165-60-0	Nitrobenzene-d5	93.2		44 - 120		93%	SPK: 100
321-60-8	2-Fluorobiphenyl	93.8		44 - 119		94%	SPK: 100
118-79-6	2,4,6-Tribromophenol	151		43 - 140		101%	SPK: 150
1718-51-0	Terphenyl-d14	101		50 - 134		101%	SPK: 100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	178000	6.875				
1146-65-2	Naphthalene-d8	702000	8.157				
15067-26-2	Acenaphthene-d10	402000	9.916				
1517-22-2	Phenanthrene-d10	683000	11.398				
1719-03-5	Chrysene-d12	437000	14.033				
1520-96-3	Perylene-d12	393000	15.504				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products