

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1539 SAS No.: Q1539 SDG NO.: Q1539
 Lab File ID: VX045238.D Date Analyzed: 03/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:58
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	103487	5.54	178691	6.76	155011	10.06
UPPER LIMIT	206974	6.043	357382	7.257	310022	10.555
LOWER LIMIT	51743.5	5.043	89345.5	6.257	77505.5	9.555
EPA SAMPLE NO.						
TAPIAL3-MW03D-031025-00-T1	66308	5.55	129212	6.76	122481	10.05
TAPFTA-MW01I-031025-00-T2	67874	5.55	131712	6.76	122865	10.05
TAP-TB-03-031025	65606	5.55	129853	6.76	121501	10.05
TAP-TB-04-031025-T2	64979	5.54	127269	6.76	118251	10.06
VX0312WBL01	62202	5.55	124010	6.76	114608	10.05
VX0312WBS01	98902	5.55	179533	6.76	158824	10.05
VX0312WBSD01	92230	5.54	161778	6.76	146828	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.