

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1539 SAS No.: Q1539 SDG NO.: Q1539
 Lab File ID: VX045238.D Date Analyzed: 03/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:58
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	69105	12.018				
UPPER LIMIT	138210	12.518				
LOWER LIMIT	34552.5	11.518				
EPA SAMPLE NO.						
TAPIAL3-MW03D-031025-00-T1	51784	12.02				
TAPFTA-MW01I-031025-00-T2	51924	12.02				
TAP-TB-03-031025	51277	12.02				
TAP-TB-04-031025-T2	46701	12.02				
VX0312WBL01	43950	12.02				
VX0312WBS01	69454	12.02				
VX0312WBSD01	67806	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.