

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q1540
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1540 **SAS No.:** Q1540
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	532	510	104	95 - 105	P	03/12/2025	10:52	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q1540
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1540 **SAS No.:** Q1540
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.9	20.0	114	80 - 120	P	03/12/2025	10:56	LB135011

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q1540
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1540 **SAS No.:** Q1540
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1260	1250	101	90 - 110	P	03/12/2025	11:29	LB135011
CCV02	Copper	1280	1250	102	90 - 110	P	03/12/2025	12:20	LB135011
CCV03	Copper	1280	1250	102	90 - 110	P	03/12/2025	13:10	LB135011
CCV04	Copper	1290	1250	103	90 - 110	P	03/12/2025	14:24	LB135011
CCV05	Copper	1260	1250	101	90 - 110	P	03/12/2025	15:25	LB135011
CCV06	Copper	1260	1250	100	90 - 110	P	03/12/2025	16:25	LB135011
CCV07	Copper	1280	1250	102	90 - 110	P	03/12/2025	17:25	LB135011
CCV08	Copper	1270	1250	102	90 - 110	P	03/12/2025	18:23	LB135011
CCV09	Copper	1280	1250	102	90 - 110	P	03/12/2025	18:49	LB135011