

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Nobis Group **SDG No.:** Q1540
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1540 **SAS No.:** Q1540
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	7.26	2.0	363	-18	22	03/12/2025	11:09	LB135011
ICSAB01	Copper	509	511	100	434	588	03/12/2025	11:14	LB135011