

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Nobis Group **level:** low **sdg no.:** Q1540
contract: NOBI03 **lab code:** CHEM **case no.:** Q1540 **sas no.:** Q1540
matrix: Water **sample id:** Q1540-01 **client id:** OU4-CF-15R-031025MS
Percent Solids for Sample: NA **Spiked ID:** Q1540-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	86 - 114	560		306		300	85	N	P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Nobis Group</u>	level:	<u>low</u>	sdg no.:	<u>Q1540</u>		
contract:	<u>NOBI03</u>	lab code:	<u>CHEM</u>	case no.:	<u>Q1540</u>	sas no.:	<u>Q1540</u>
matrix:	<u>Water</u>	sample id:	<u>Q1540-01</u>	client id:	<u>OU4-CF-15R-031025MSD</u>		
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>Q1540-01MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>		

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	86 - 114	559		306		300	85	N	P