

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q1540
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q1540 **SAS No.:** Q1540

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167099BS Copper	ug/L	300	305		102	86 - 114	P