DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name :		Alliance Technical Group LLC	Client :	Sciacca General Contractors, LLC					
Project Location : Project Number :									
Laboratory Sample ID(s) : Q1542			Sampling Date(s):	3/10/2025					
List DKQP Methods Used (e.g., 8260,8270, et Cetra) 8015D,8260D,NJEPH,SOP									
1	specified QA/Q explain any crite	cical method referenced in this labo C performance criteria followed, in eria falling outside of acceptable g Known Quality performance stanc	cluding the requirement to uidelines, as specified in the		V	Yes		No	
1A	Were the method specified handling, preservation, and holding time requirements met?					Yes		No	
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)					Yes		No	□ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?				$\overline{\mathbf{V}}$	Yes		No	
3	Were samples received at an appropriate temperature (4±2° C)?				V	Yes		No	□ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?					Yes		No	
5		ng limits specified or referenced on to the laboratory prior to sample re	-		V	Yes		No	
	b)Were these re	eporting limits met?			$\overline{\mathbf{V}}$	Yes		No	□ N/A
6	results reported	ical method referenced in this labor d for all constituents identified in the e DKQP documents and/or site-sp	e method-specific analyte lists		$\overline{\mathbf{V}}$	Yes		No	
7	Are project-spe	cific matrix spikes and/or laborator	y duplicates included in this dat	a set?	V	Yes		No	

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."