

Continuing Calibration Report for SequenceID : FE031225AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE052771.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C12-C16 | 5691567.000  | 40.000  | 6.791  | 10.243 | 142289.175 | 152098.489 | 6.449  |
| Aliphatic C9-C12  | 5804698.000  | 60.000  | 3.148  | 6.790  | 96744.967  | 155948.436 | 37.963 |
| Aliphatic C16-C21 | 8107573.000  | 60.000  | 10.244 | 13.621 | 135126.217 | 146548.085 | 7.794  |
| Aliphatic C21-C28 | 10769149.000 | 80.000  | 13.622 | 17.293 | 134614.363 | 142451.851 | 5.502  |
| Aliphatic C28-C40 | 15618020.000 | 120.000 | 17.294 | 22.210 | 130150.167 | 127589.108 | -2.007 |
| Aliphatic EPH     | 45991007.000 | 360.000 | 3.148  | 22.210 | 127752.797 | 141501.588 | 9.716  |