

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: SCIA01
 ProjectID: 248 Union St., Lodi
 Lab Code: CHEM Case No.: Q1542 SAS No.: Q1542 SDG No.: Q1542

Calibration Sequence : FF030325		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	205033948	120608	FF015569.D
850	98410584	115777	FF015570.D
340	39379290	115821	FF015571.D
170	20344477	119673	FF015572.D
85	10098758	118809	FF015573.D
AVG RF : 118138		% RSD : 1.886	AVG RT : 15.0244