

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030425

Review By	Mahesh Dadoda	Review On	3/5/2025 12:14:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/5/2025 12:18:08 PM		
SubDirectory	VY030425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133216
Initial Calibration Stds	VP133207,VP133208,VP133209,VP133210,VP133211,VP133212
CCC	VP133214,VP133215
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133213
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021395.D	04 Mar 2025 08:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021396.D	04 Mar 2025 09:23	not used	SY/MD	Not Ok
3	VSTDICC010	VSTDICC010	VY021397.D	04 Mar 2025 09:46		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021398.D	04 Mar 2025 10:09	Comp.#11 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021399.D	04 Mar 2025 10:30	Comp.#95 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021400.D	04 Mar 2025 11:06		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021401.D	04 Mar 2025 11:29	Method fail for comp.#13	SY/MD	Ok,M
8	VIBLK	VIBLK	VY021402.D	04 Mar 2025 11:52		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY021403.D	04 Mar 2025 12:15		SY/MD	Ok,M
10	VSTDICV050	ICVVY030425	VY021404.D	04 Mar 2025 13:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY031325

Review By	Review On				
Supervise By	Supervise On				
SubDirectory	VY031325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133260				
CCC	VP133261,VP133262				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021515.D	13 Mar 2025 09:58		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021516.D	13 Mar 2025 10:31		SY/MD	Ok,NR
3	VY0313SBL01	VY0313SBL01	VY021517.D	13 Mar 2025 11:27		SY/MD	Ok
4	VY0313SBS01	VY0313SBS01	VY021518.D	13 Mar 2025 12:13		SY/MD	Ok,NR
5	VY0313SBSD01	VY0313SBSD01	VY021519.D	13 Mar 2025 12:35		SY/MD	Ok,NR
6	Q1545-02	BU-02-031125	VY021520.D	13 Mar 2025 13:11	vial-B not purge	SY/MD	Not Ok
7	Q1553-01	RT2321	VY021521.D	13 Mar 2025 13:34	Vial-A Internal standard Fail	SY/MD	ReRun
8	Q1542-02	VOC	VY021522.D	13 Mar 2025 13:58	vial-B ISTD Fail	SY/MD	Ok,NR
9	Q1556-01	OR-02-031225	VY021523.D	13 Mar 2025 14:21	Vial-A not purged	SY/MD	Not Ok
10	Q1552-02	NB-418-JB-VOC-01	VY021524.D	13 Mar 2025 14:44	vial-A	SY/MD	Ok
11	Q1549-01	72-11978	VY021525.D	13 Mar 2025 15:08	vial-A ISTD Fail	SY/MD	Ok
12	Q1549-01RE	72-11978RE	VY021526.D	13 Mar 2025 15:41	vial-B not purge	SY/MD	Not Ok
13	Q1556-01	OR-02-031225	VY021527.D	13 Mar 2025 16:05	vial-B	SY/MD	Ok
14	VSTDCCC050	VSTDCCC050EC	VY021528.D	13 Mar 2025 16:30		SY/MD	Ok,NR

M : Manual Integration