

Surrogate Summary

SDG No.: Q1542

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1542-02	VOC	1,2-Dichloroethane-d4	50	55.1	110	70 (63)	130 (155)
		Dibromofluoromethane	50	53.0	106	70 (70)	130 (134)
		Toluene-d8	50	47.7	95	70 (74)	130 (123)
		4-Bromofluorobenzene	50	37.6	75	70 (38)	130 (136)
VY0313SBL01	VY0313SBL01	1,2-Dichloroethane-d4	50	59.2	118	70 (63)	130 (155)
		Dibromofluoromethane	50	51.6	103	70 (70)	130 (134)
		Toluene-d8	50	49.0	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	43.9	88	70 (38)	130 (136)
VY0313SBS01	VY0313SBS01	1,2-Dichloroethane-d4	50	54.2	108	70 (63)	130 (155)
		Dibromofluoromethane	50	53.3	107	70 (70)	130 (134)
		Toluene-d8	50	53.7	107	70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.5	105	70 (38)	130 (136)
VY0313SBSD01	VY0313SBSD01	1,2-Dichloroethane-d4	50	58.5	117	70 (63)	130 (155)
		Dibromofluoromethane	50	57.6	115	70 (70)	130 (134)
		Toluene-d8	50	56.6	113	70 (74)	130 (123)
		4-Bromofluorobenzene	50	56.8	114	70 (38)	130 (136)

