

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCIA01
 Lab Code: CHEM Case No.: Q1542 SAS No.: Q1542 SDG NO.: Q1542
 Lab File ID: VY021516.D Date Analyzed: 03/13/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	202933	7.71	311585	8.62	288548	11.41
UPPER LIMIT	405866	8.207	623170	9.116	577096	11.914
LOWER LIMIT	101467	7.207	155793	8.116	144274	10.914
EPA SAMPLE NO.						
VOC	135626	7.71	228055	8.62	188696	11.41
VY0313SBL01	239160	7.71	443031	8.62	398087	11.41
VY0313SBS01	202139	7.71	319039	8.62	282466	11.42
VY0313SBSD01	190559	7.71	296797	8.62	261921	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.