

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCIA01
 Lab Code: CHEM Case No.: Q1542 SAS No.: Q1542 SDG NO.: Q1542
 Lab File ID: VY021516.D Date Analyzed: 03/13/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	154472	13.346				
UPPER LIMIT	308944	13.846				
LOWER LIMIT	77236	12.846				
EPA SAMPLE NO.						
VOC	65866 *	13.35				
VY0313SBL01	156172	13.35				
VY0313SBS01	145688	13.35				
VY0313SBSD01	136724	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.