

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ATCE02

Lab Code: CHEM **Case No.:** Q1544 **SAS No.:** Q1544 **SDG NO.:** Q1544

Instrument ID: ECD_O **Date(s) Analyzed:** 02/20/2025 02/21/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.91	3.81	4.01	131879000
		2	4.00	3.90	4.10	98701000
		3	4.07	3.97	4.17	269838000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.07	3.97	4.17	208928000
		2	4.57	4.47	4.67	116083000
		3	4.81	4.71	4.91	201936000
		4	4.99	4.89	5.09	111220000
		5	5.03	4.93	5.13	81668600
Aroclor-1262	500	1	6.85	6.75	6.95	665842000
		2	7.35	7.25	7.45	1129720000
		3	7.63	7.53	7.73	448264000
		4	7.70	7.60	7.80	829210000
		5	8.19	8.09	8.29	373330000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ATCE02

Lab Code: CHEM **Case No.:** Q1544 **SAS No.:** Q1544 **SDG NO.:** Q1544

Instrument ID: ECD_O **Date(s) Analyzed:** 02/20/2025 02/21/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.91	3.81	4.01	68502400
		2	3.99	3.89	4.09	51339800
		3	4.07	3.97	4.17	145428000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.07	3.97	4.17	111462000
		2	4.80	4.70	4.90	103317000
		3	4.97	4.87	5.07	56771800
		4	5.06	4.96	5.16	54894600
		5	5.23	5.13	5.33	58805000
Aroclor-1262	500	1	6.81	6.71	6.91	317212000
		2	7.31	7.21	7.41	498620000
		3	7.60	7.50	7.70	196255000
		4	7.66	7.56	7.76	354890000
		5	8.15	8.05	8.25	143138000