

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1544 SAS No.: Q1544 SDG NO.: Q1544

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PO109817.D Time Analyzed: 12:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.786	4.691	4.891	553.870	500.000	10.8
Aroclor-1016-2	4.805	4.711	4.911	551.680	500.000	10.3
Aroclor-1016-3	4.861	4.767	4.967	554.170	500.000	10.8
Aroclor-1016-4	4.981	4.888	5.088	554.930	500.000	11.0
Aroclor-1016-5	5.239	5.145	5.345	541.140	500.000	8.2
Aroclor-1260-1	6.281	6.188	6.388	526.350	500.000	5.3
Aroclor-1260-2	6.470	6.377	6.577	526.430	500.000	5.3
Aroclor-1260-3	6.838	6.746	6.946	532.650	500.000	6.5
Aroclor-1260-4	7.098	7.005	7.205	526.890	500.000	5.4
Aroclor-1260-5	7.341	7.247	7.447	526.460	500.000	5.3
Decachlorobiphenyl	8.748	8.657	8.857	45.040	50.000	-9.9
Tetrachloro-m-xylene	3.692	3.598	3.798	52.800	50.000	5.6

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1544 SAS No.: Q1544 SDG NO.: Q1544

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PO109817.D Time Analyzed: 12:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.772	4.678	4.878	550.620	500.000	10.1
Aroclor-1016-2	4.792	4.697	4.897	557.960	500.000	11.6
Aroclor-1016-3	4.967	4.872	5.072	559.560	500.000	11.9
Aroclor-1016-4	5.009	4.914	5.114	542.370	500.000	8.5
Aroclor-1016-5	5.222	5.127	5.327	537.850	500.000	7.6
Aroclor-1260-1	6.254	6.161	6.361	519.200	500.000	3.8
Aroclor-1260-2	6.442	6.348	6.548	523.500	500.000	4.7
Aroclor-1260-3	6.595	6.502	6.702	564.690	500.000	12.9
Aroclor-1260-4	7.067	6.974	7.174	517.950	500.000	3.6
Aroclor-1260-5	7.307	7.214	7.414	525.390	500.000	5.1
Decachlorobiphenyl	8.698	8.607	8.807	44.410	50.000	-11.2
Tetrachloro-m-xylene	3.690	3.594	3.794	52.950	50.000	5.9