

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1544 SAS No.: Q1544 SDG NO.: Q1544

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PO109826.D Time Analyzed: 16:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.788	4.691	4.891	513.000	500.000	2.6
Aroclor-1016-2	4.807	4.711	4.911	524.660	500.000	4.9
Aroclor-1016-3	4.864	4.767	4.967	523.430	500.000	4.7
Aroclor-1016-4	4.984	4.888	5.088	522.390	500.000	4.5
Aroclor-1016-5	5.242	5.145	5.345	511.860	500.000	2.4
Aroclor-1260-1	6.283	6.188	6.388	498.930	500.000	-0.2
Aroclor-1260-2	6.472	6.377	6.577	502.080	500.000	0.4
Aroclor-1260-3	6.840	6.746	6.946	505.260	500.000	1.1
Aroclor-1260-4	7.101	7.005	7.205	501.060	500.000	0.2
Aroclor-1260-5	7.342	7.247	7.447	502.040	500.000	0.4
Decachlorobiphenyl	8.748	8.657	8.857	43.150	50.000	-13.7
Tetrachloro-m-xylene	3.694	3.598	3.798	49.280	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1544 SAS No.: Q1544 SDG NO.: Q1544

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: AR1660CCC500 Data File : PO109826.D Time Analyzed: 16:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.774	4.678	4.878	528.970	500.000	5.8
Aroclor-1016-2	4.794	4.697	4.897	547.240	500.000	9.4
Aroclor-1016-3	4.969	4.872	5.072	541.490	500.000	8.3
Aroclor-1016-4	5.011	4.914	5.114	524.310	500.000	4.9
Aroclor-1016-5	5.224	5.127	5.327	524.040	500.000	4.8
Aroclor-1260-1	6.257	6.161	6.361	505.170	500.000	1.0
Aroclor-1260-2	6.444	6.348	6.548	506.900	500.000	1.4
Aroclor-1260-3	6.597	6.502	6.702	544.080	500.000	8.8
Aroclor-1260-4	7.068	6.974	7.174	498.410	500.000	-0.3
Aroclor-1260-5	7.309	7.214	7.414	503.000	500.000	0.6
Decachlorobiphenyl	8.700	8.607	8.807	41.670	50.000	-16.7
Tetrachloro-m-xylene	3.692	3.594	3.794	51.210	50.000	2.4