

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135014

Slope : 98.9

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/13/2025	08:50
2	CAL2	1	Water	NA	NA	20.3	7.00	03/13/2025	08:51
3	CAL3	1	Water	NA	NA	20.3	10.02	03/13/2025	08:52
4	ICV	1	Water	NA	NA	20.3	7.01	03/13/2025	09:00
5	CCV1	1	Water	NA	NA	20.3	2.01	03/13/2025	09:03
6	Q1548-01	1	Solid	20.02	20	21.4	6.29	03/13/2025	09:14
7	Q1548-01DUP	1	Solid	20.03	20	21.5	6.30	03/13/2025	09:15
8	Q1548-02	1	Solid	20.02	20	21.3	5.63	03/13/2025	09:25
9	Q1548-03	1	Solid	20.03	20	20.6	4.57	03/13/2025	09:33
10	CCV2	1	Water	NA	NA	20.3	12.02	03/13/2025	09:35

WORKLIST(Hardcopy Internal Chain)

135014

WorkList Name : PH Q1548

WorkList ID : 188248

Department : Wet-Chemistry

Date : 03-13-2025 08:31:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1548-01	TAPLPR-SED12-031125-00-T2	Solid	pH	Cool 4 deg C	WEST04	I41	03/11/2025	9045D
Q1548-02	TAPLPR-SED13-031125-00-T2	Solid	pH	Cool 4 deg C	WEST04	I41	03/11/2025	9045D
Q1548-03	TAPIAL2-SED03-031125-00-T2	Solid	pH	Cool 4 deg C	WEST04	I41	03/11/2025	9045D

Date/Time 03/13/25 08:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 03/13/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

12:00

[Signature]

[Signature]