

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1548

Project: Fort Meade MD Tipton Airfield Parcel RI - 0111169

RunNo.: LB135014

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	03/13/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	03/13/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	03/13/2025

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1548

Project: Fort Meade MD Tipton Airfield Parcel RI - 0111169

RunNo.: LB135101

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	1030	1000	103	90-110	03/14/2025
Sample ID:	CCV1						
TOC		mg/L	1030	1000	103	90-110	03/20/2025
Sample ID:	CCV2						
TOC		mg/L	1050	1000	105	90-110	03/20/2025