

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1548

Project: Fort Meade MD Tipton Airfield Parcel RI - 0111169

RunNo.: LB135101

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	03/14/2025
Sample ID: CCB1 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	03/20/2025
Sample ID: CCB2 TOC	mg/L	< 125.0000	125.0000	U	22.3	250	03/20/2025