

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1548
Project:	Fort Meade MD Tipton Airfield Parcel RI - 0111169	Sample ID:	Q1548-02
Client ID:	TAPLPR-SED13-031125-00-T2MS	Percent Solids for Spike Sample:	55.5

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	16300	OR	15500	OR	1000	1	80		03/20/2025

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1548
Project:	Fort Meade MD Tipton Airfield Parcel RI - 0111169	Sample ID:	Q1548-02
Client ID:	TAPLPR-SED13-031125-00-T2MSD	Percent Solids for Spike Sample:	55.5

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	16200	OR	15500	OR	1000	1	70	*	03/20/2025