

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1548
Project: Fort Meade MD Tipton Airfield Parcel RI - 0111169	Sample ID: Q1548-01
Client ID: TAPLPR-SED12-031125-00-T2DUP	Percent Solids for Spike Sample: 84.6

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	6.29		6.30		1	0.16		03/13/2025

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1548
Project: Fort Meade MD Tipton Airfield Parcel RI - 0111169	Sample ID: Q1548-02
Client ID: TAPLPR-SED13-031125-00-T2MSD	Percent Solids for Spike Sample: 55.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TOC	mg/Kg	+/-20	16300	OR	16200	OR	1	1		03/20/2025