

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1557 SAS No.: Q1557 SDG No.: Q1557
 Instrument ID: BNA_N Calibration Date/Time: 03/14/2025 10:09
 Lab File ID: BN036601.D Init. Calib. Date(s): 03/10/2025 03/10/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:42 15:19
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.593		-0.3	20.0
Fluoranthene-d10	1.025	1.108		8.1	20.0
2-Fluorophenol	0.932	0.929		-0.3	20.0
Phenol-d6	1.152	1.129		-2.0	20.0
Nitrobenzene-d5	0.435	0.423		-2.8	20.0
2-Fluorobiphenyl	2.327	2.416		3.8	20.0
2,4,6-Tribromophenol	0.182	0.184		1.1	20.0
Terphenyl-d14	0.958	0.974		1.7	20.0
1,4-Dioxane	0.444	0.525		18.2	20.0

All other compounds must meet a minimum RRF of 0.010.