

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031425\
 Data File : BN036631.D
 Acq On : 15 Mar 2025 06:04
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 15 06:29:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.450	-12.5	98	0.00
3	n-Nitrosodimethylamine	0.400	0.433	-8.2	101	0.00
4 S	2-Fluorophenol	0.400	0.410	-2.5	94	0.00
5 S	Phenol-d6	0.400	0.399	0.3	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	104	0.00
8 S	Nitrobenzene-d5	0.400	0.378	5.5	103	-0.01
9	Naphthalene	0.400	0.405	-1.3	103	-0.01
10	Hexachlorobutadiene	0.400	0.414	-3.5	101	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.397	0.8	101	-0.01
12	2-Methylnaphthalene	0.400	0.406	-1.5	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.398	0.5	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	101	0.00
16	Acenaphthylene	0.400	0.409	-2.2	103	0.00
17	Acenaphthene	0.400	0.418	-4.5	104	-0.01
18	Fluorene	0.400	0.418	-4.5	102	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.427	-6.7	97	0.00
21	4-Bromophenyl-phenylether	0.400	0.426	-6.5	102	0.00
22	Hexachlorobenzene	0.400	0.427	-6.7	100	0.00
23	Atrazine	0.400	0.420	-5.0	103	0.00
24	Pentachlorophenol	0.400	0.363	9.3	95	-0.01
25	Phenanthrene	0.400	0.425	-6.2	102	0.00
26	Anthracene	0.400	0.418	-4.5	103	-0.01
27 SURR	Fluoranthene-d10	0.400	0.422	-5.5	101	0.00
28	Fluoranthene	0.400	0.426	-6.5	103	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.453	-13.2	102	0.00
31 S	Terphenyl-d14	0.400	0.434	-8.5	99	0.00
32	Benzo(a)anthracene	0.400	0.411	-2.7	98	0.00
33	Chrysene	0.400	0.449	-12.2	104	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.474	-18.5	110	0.00
35 I	Perylene-d12	0.400	0.400	0.0	95	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.436	-9.0	97	-0.02
37	Benzo(b)fluoranthene	0.400	0.457	-14.2	102	0.00
38	Benzo(k)fluoranthene	0.400	0.450	-12.5	101	0.00
39 C	Benzo(a)pyrene	0.400	0.451	-12.7	101	0.00
40	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	94	-0.01
41	Benzo(g,h,i)perylene	0.400	0.441	-10.2	97	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4