



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1557

Client: AECOM Technical Services, Inc.

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1557-04MS	Client Sample ID:	BPOW6-9-20250312MS					DataFile:	BN036611.D		
1,4-Dioxane	0.4	0	0.22	ug/L	55	*			70	130	