

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1557 SAS No.: Q1557 SDG NO.: Q1557

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/15/2025

Lab File ID: BN036633.D Time Analyzed: 08:35

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2056	7.717	5167	10.51	3028	14.36
UPPER LIMIT	4112	8.217	10334	11.009	6056	14.856
LOWER LIMIT	1028	7.217	2583.5	10.009	1514	13.856
EPA SAMPLE NO.						
01 PB167128BL	2323	7.72	4860	10.52	2639	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.