

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1557 SAS No.: Q1557 SDG NO.: Q1557

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/15/2025

Lab File ID: BN036633.D Time Analyzed: 08:35

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5852	17.099	3964	21.295	3322	23.543
UPPER LIMIT	11704	17.599	7928	21.795	6644	24.043
LOWER LIMIT	2926	16.599	1982	20.795	1661	23.043
EPA SAMPLE NO.						
PB167128BL	4474	17.12	2905	21.30	2592	23.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.