

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1557 SAS No.: Q1557 SDG No.: Q1557

Instrument ID: MSVOA_X Calibration Date/Time: 03/18/2025 20:16

Lab File ID: VX045338.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.613		-2.54	50
Chloromethane	0.770	0.675	0.1	-12.34	50
Vinyl Chloride	0.764	0.692		-9.42	50
Bromomethane	0.300	0.300		0	50
Chloroethane	0.352	0.317		-9.94	50
Trichlorofluoromethane	1.012	1.031		1.88	50
1,1,2-Trichlorotrifluoroethane	0.581	0.667		14.8	50
1,1-Dichloroethene	0.614	0.645		5.05	50
Acetone	0.369	0.378		2.44	50
Carbon Disulfide	1.636	1.376		-15.89	50
Methyl tert-butyl Ether	2.061	2.347		13.88	50
Methyl Acetate	0.888	1.198		34.91	50
Methylene Chloride	0.731	0.762		4.24	50
trans-1,2-Dichloroethene	0.607	0.652		7.41	50
1,1-Dichloroethane	1.247	1.406	0.1	12.75	50
Cyclohexane	1.082	1.119		3.42	50
2-Butanone	0.555	0.589		6.13	50
Carbon Tetrachloride	0.465	0.533		14.62	50
cis-1,2-Dichloroethene	0.749	0.827		10.41	50
Bromochloromethane	0.594	0.656		10.44	50
Chloroform	1.239	1.421		14.69	50
1,1,1-Trichloroethane	1.000	1.177		17.7	50
Methylcyclohexane	0.549	0.616		12.2	50
Benzene	1.448	1.576		8.84	50
1,2-Dichloroethane	0.521	0.625		19.96	50
Trichloroethene	0.340	0.369		8.53	50
1,2-Dichloropropane	0.372	0.418		12.37	50
Bromodichloromethane	0.516	0.602		16.67	50
4-Methyl-2-Pentanone	0.587	0.651		10.9	50
Toluene	0.849	0.954		12.37	50
t-1,3-Dichloropropene	0.431	0.523		21.35	50
cis-1,3-Dichloropropene	0.503	0.601		19.48	50
1,1,2-Trichloroethane	0.350	0.385		10	50
2-Hexanone	0.425	0.472		11.06	50
Dibromochloromethane	0.366	0.413		12.84	50
1,2-Dibromoethane	0.349	0.387		10.89	50
Tetrachloroethene	0.320	0.359		12.19	50
Chlorobenzene	1.058	1.176	0.3	11.15	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1557 SAS No.: Q1557 SDG No.: Q1557

Instrument ID: MSVOA_X Calibration Date/Time: 03/18/2025 20:16

Lab File ID: VX045338.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.137		15.95	50
m/p-Xylenes	0.675	0.777		15.11	50
o-Xylene	0.681	0.758		11.31	50
Styrene	1.101	1.274		15.71	50
Bromoform	0.266	0.297	0.1	11.65	50
Isopropylbenzene	3.903	4.486		14.94	50
1,1,2,2-Tetrachloroethane	1.432	1.566	0.3	9.36	50
1,3-Dichlorobenzene	1.643	1.847		12.42	50
1,4-Dichlorobenzene	1.662	1.846		11.07	50
1,2-Dichlorobenzene	1.648	1.889		14.62	50
1,2-Dibromo-3-Chloropropane	0.279	0.331		18.64	50
1,2,4-Trichlorobenzene	0.938	1.113		18.66	50
1,2,3-Trichlorobenzene	0.990	1.174		18.59	50
1,2-Dichloroethane-d4	0.795	0.973		22.39	50
Dibromofluoromethane	0.334	0.396		18.56	50
Toluene-d8	1.212	1.395		15.1	50
4-Bromofluorobenzene	0.402	0.479		19.15	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.