

Surrogate Summary

SDG No.: Q1557

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1557-01	BPOW6-7-20250312	1,2-Dichloroethane-d4	50	56.4	113	81	118
		Dibromofluoromethane	50	53.4	107	80	119
		Toluene-d8	50	51.8	104	89	112
		4-Bromofluorobenzene	50	54.3	109	85	114
Q1557-02	BPOW6-8-20250312	1,2-Dichloroethane-d4	50	55.3	111	81	118
		Dibromofluoromethane	50	52.9	106	80	119
		Toluene-d8	50	52.1	104	89	112
		4-Bromofluorobenzene	50	49.7	99	85	114
Q1557-03	BPOW6-9-20250312	1,2-Dichloroethane-d4	50	57.3	115	81	118
		Dibromofluoromethane	50	51.8	104	80	119
		Toluene-d8	50	52.3	105	89	112
		4-Bromofluorobenzene	50	55.1	110	85	114
Q1557-04MS	BPOW6-9-20250312MS	1,2-Dichloroethane-d4	50	56.9	114	81	118
		Dibromofluoromethane	50	53.8	108	80	119
		Toluene-d8	50	52.2	104	89	112
		4-Bromofluorobenzene	50	56.3	112	85	114
Q1557-05MSD	BPOW6-9-20250312MSD	1,2-Dichloroethane-d4	50	52.2	104	81	118
		Dibromofluoromethane	50	49.6	99	80	119
		Toluene-d8	50	48.6	97	89	112
		4-Bromofluorobenzene	50	54.2	108	85	114
Q1557-06	BPOW6-10-20250312	1,2-Dichloroethane-d4	50	56.6	113	81	118
		Dibromofluoromethane	50	51.5	103	80	119
		Toluene-d8	50	51.6	103	89	112
		4-Bromofluorobenzene	50	53.7	107	85	114
Q1557-07	BPOW6-11-20250312	1,2-Dichloroethane-d4	50	56.5	113	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	50.7	101	85	114
Q1557-08	DUP04-20250312	1,2-Dichloroethane-d4	50	57.5	115	81	118
		Dibromofluoromethane	50	52.2	104	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	55.3	111	85	114
VX0318WBL01	VX0318WBL01	1,2-Dichloroethane-d4	50	56.3	113	81	118
		Dibromofluoromethane	50	54.0	108	80	119
		Toluene-d8	50	52.3	105	89	112
		4-Bromofluorobenzene	50	52.2	104	85	114
VX0318WBS01	VX0318WBS01	1,2-Dichloroethane-d4	50	53.2	106	81	118
		Dibromofluoromethane	50	52.9	106	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	54.1	108	85	114