



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**AECOM Technical Services, Inc.**

**Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258**

**Project # N/A**

**Chemtech Project # Q1559**

**Test Name: VOCMS Group1**

### **A. Number of Samples and Date of Receipt:**

22 Water samples were received on 03/12/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group1 was based on method 8260D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VX045312.D met the requirements except for Carbon Disulfide which is our target compound but failing marginally low, therefore no corrective action taken.

The Continuous Calibration File ID VX045347.D met the requirements except for Methyl Acetate and t-1,3-Dichloropropene are failing high but no positive hit in associate sample therefore no corrective action taken.

The Tuning criteria met requirements.

Samples RE115D2-20250310 was direct diluted due to past history of this sample containing high amounts of Trichloroethene.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

Samples RE115D1-20250310, RE134D3-20250310, RE134D4-20250310, RE125D3-20250310, DUP03-20250310, RW4-RE137-20250311, RE120D3-20250312 and RE120D2-20250312 were diluted due to high concentrations.

**E. Additional Comments:**

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.  
The not QT review data is reported in the Miscellaneous.

The Sample #RE115D2-20250310 and RE134D1-20250310 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_