

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1559

MATRIX: Water

METHOD: 8270-Modified/3510

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5. GC/MS Calibration Requirements.			✓
The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			
6. Blank Contamination - If yes, list compounds and concentrations in each blank:			✓
7. Surrogate Recoveries Meet Criteria			✓

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

The Surrogate recoveries met the acceptable criteria except for BPOW6-9-20250312MS [Terphenyl-d14 - 163%], BPOW6-9-20250312MSD [Terphenyl-d14 - 172%], RE115D1-20250310 [Terphenyl-d14 - 151%], RE115D1-20250310DL [2-Fluorobiphenyl - 126%, Terphenyl-d14 - 180%], RE115D2-20250310 [2-Fluorobiphenyl - 108%, Terphenyl-d14 - 140%], RE134D1-20250310 [2-Fluorobiphenyl - 114%], RE134D3-20250310 [2-Fluorobiphenyl - 111%, Terphenyl-d14 - 134%], RE134D4-20250310 [2-Fluorobiphenyl - 109%, Terphenyl-d14 - 138%] and RE120D3-20250312DL [Terphenyl-d14 - 139%], failure surrogates are not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken also RE115D2-20250310DL [2-Fluorobiphenyl - 154%, 2-Methylnaphthalene-d10 - 159%, Fluoranthene-d10 - 200%, Nitrobenzene-d5 - 144%, Terphenyl-d14 - 210%], RE134D3-20250310DL [2-Fluorobiphenyl - 160%, Fluoranthene-d10 - 184%, Nitrobenzene-d5 - 135%, Terphenyl-d14 - 211%], but these samples were required dilution as well due to high concentration, therefore original and Dilution analysis were reported and no further corrective action taken.

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

		NA	NO	YES
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The MS {Q1557-04MS} with File ID: BN036611.D recoveries met the requirements for all compounds except for 1,4-Dioxane[55%] due to matrix interference. The MSD {Q1557-05MSD} with File ID: BN036612.D recoveries met the acceptable requirements except for 1,4-Dioxane[58%] due to matrix interference.			
	The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

Samples RE115D1-20250310, RE115D2-20250310, RE134D3-20250310, DUP03-20250310 and RE120D3-20250312 were diluted due to high concentrations.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date