



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1559

Client: AECOM Technical Services, Inc.

Analytical Method: SW8270-Modified

| Parameter                  | Spike | Sample Result                         | Result | Units | Rec | Rec Qual | RPD | RPD Qual             | Low | Limits High | RPD |
|----------------------------|-------|---------------------------------------|--------|-------|-----|----------|-----|----------------------|-----|-------------|-----|
| Lab Sample ID: Q1557-05MSD |       | Client Sample ID: BPOW6-9-20250312MSD |        |       |     |          |     | DataFile: BN036612.D |     |             |     |
| 1,4-Dioxane                | 0.4   | 0                                     | 0.23   | ug/L  | 58  | *        | 5   |                      | 70  | 130         | 20  |