

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/15/2025

Lab File ID: BN036633.D Time Analyzed: 08:35

Instrument ID: BNA\_N GC Column: ZB-GR ID: 0.25 (mm)

|    |                      | IS4 (PHN)<br>AREA # | RT #   | IS5 (CRY)<br>AREA # | RT #   | IS6 (PRY)<br>AREA # | RT #   |
|----|----------------------|---------------------|--------|---------------------|--------|---------------------|--------|
|    | 12 HOUR STD          | 5852                | 17.099 | 3964                | 21.295 | 3322                | 23.543 |
|    | UPPER LIMIT          | 11704               | 17.599 | 7928                | 21.795 | 6644                | 24.043 |
|    | LOWER LIMIT          | 2926                | 16.599 | 1982                | 20.795 | 1661                | 23.043 |
|    | EPA SAMPLE NO.       |                     |        |                     |        |                     |        |
| 01 | PB167128BL           | 4474                | 17.12  | 2905                | 21.30  | 2592                | 23.56  |
| 02 | PB167131BS           | 5517                | 17.10  | 3441                | 21.29  | 2899                | 23.54  |
| 03 | PB167131BSD          | 4741                | 17.10  | 2695                | 21.30  | 2381                | 23.55  |
| 04 | TT149S1-20250310     | 4466                | 17.10  | 3384                | 21.30  | 2800                | 23.54  |
| 05 | RE125D3-20250310     | 6919                | 17.10  | 4660                | 21.30  | 3624                | 23.55  |
| 06 | DUP02-20250310       | 6912                | 17.10  | 4784                | 21.30  | 4070                | 23.54  |
| 07 | DUP03-20250310       | 6238                | 17.11  | 4680                | 21.30  | 3834                | 23.55  |
| 08 | RW10-MW01D-20250311  | 4969                | 17.10  | 4107                | 21.30  | 3278                | 23.54  |
| 09 | RW10-MW01S-20250311  | 5515                | 17.10  | 3873                | 21.29  | 3147                | 23.55  |
| 10 | TT158S1-20250311     | 5718                | 17.10  | 4201                | 21.30  | 3441                | 23.54  |
| 11 | TT158I1-20250311     | 5804                | 17.10  | 4480                | 21.29  | 3470                | 23.54  |
| 12 | RW10A-MW01S-20250311 | 5682                | 17.10  | 3838                | 21.30  | 3126                | 23.55  |
| 13 | RW10A-MW01I-20250311 | 6923                | 17.10  | 4981                | 21.29  | 3956                | 23.54  |
| 14 | RW4-RE137-20250311   | 6270                | 17.10  | 4242                | 21.29  | 3367                | 23.55  |
| 15 | RE104D1-20250312     | 5723                | 17.10  | 4058                | 21.30  | 3280                | 23.55  |

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/15/2025

Lab File ID: BN036633.D Time Analyzed: 08:35

Instrument ID: BNA\_N GC Column: ZB-GR ID: 0.25 (mm)

|                | IS4 (PHN)<br>AREA # | RT #   | IS5 (CRY)<br>AREA # | RT #   | IS6 (PRY)<br>AREA # | RT #   |
|----------------|---------------------|--------|---------------------|--------|---------------------|--------|
| 12 HOUR STD    | 5852                | 17.099 | 3964                | 21.295 | 3322                | 23.543 |
| UPPER LIMIT    | 11704               | 17.599 | 7928                | 21.795 | 6644                | 24.043 |
| LOWER LIMIT    | 2926                | 16.599 | 1982                | 20.795 | 1661                | 23.043 |
| EPA SAMPLE NO. |                     |        |                     |        |                     |        |

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.