

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1559

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045315.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0318WBS01	Dichlorodifluoromethane	20	18.6	ug/L	93			32	152	
	Chloromethane	20	16.7	ug/L	84			50	139	
	Vinyl chloride	20	15.9	ug/L	79			58	137	
	Bromomethane	20	18.2	ug/L	91			53	141	
	Chloroethane	20	20.3	ug/L	102			60	138	
	Trichlorofluoromethane	20	18.6	ug/L	93			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.8	ug/L	104			70	136	
	1,1-Dichloroethene	20	18.5	ug/L	93			71	131	
	Acetone	100	99.4	ug/L	99			39	160	
	Carbon disulfide	20	14.6	ug/L	73			64	133	
	Methyl tert-butyl Ether	20	19.8	ug/L	99			71	124	
	Methyl Acetate	20	24.5	ug/L	123			56	136	
	Methylene Chloride	20	18.8	ug/L	94			74	124	
	trans-1,2-Dichloroethene	20	18.7	ug/L	94			75	124	
	1,1-Dichloroethane	20	19.7	ug/L	99			77	125	
	Cyclohexane	20	18.6	ug/L	93			71	130	
	2-Butanone	100	100	ug/L	100			56	143	
	Carbon Tetrachloride	20	19.7	ug/L	99			72	136	
	cis-1,2-Dichloroethene	20	19.6	ug/L	98			78	123	
	Bromochloromethane	20	21.8	ug/L	109			78	123	
	Chloroform	20	20.5	ug/L	103			79	124	
	1,1,1-Trichloroethane	20	20.0	ug/L	100			74	131	
	Methylcyclohexane	20	19.8	ug/L	99			72	132	
	Benzene	20	19.7	ug/L	99			79	120	
	1,2-Dichloroethane	20	21.3	ug/L	106			73	128	
	Trichloroethene	20	18.9	ug/L	95			79	123	
	1,2-Dichloropropane	20	19.8	ug/L	99			78	122	
	Bromodichloromethane	20	20.6	ug/L	103			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	20.5	ug/L	103			80	121	
	t-1,3-Dichloropropene	20	20.4	ug/L	102			73	127	
	cis-1,3-Dichloropropene	20	21.2	ug/L	106			75	124	
	1,1,2-Trichloroethane	20	20.5	ug/L	103			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	20.3	ug/L	102			74	126	
	1,2-Dibromoethane	20	20.4	ug/L	102			77	121	
	Tetrachloroethene	20	20.0	ug/L	100			74	129	
	Chlorobenzene	20	19.9	ug/L	100			82	118	
	Ethyl Benzene	20	20.3	ug/L	102			79	121	
	m/p-Xylenes	40	41.2	ug/L	103			80	121	
	o-Xylene	20	20.1	ug/L	101			78	122	
	Styrene	20	20.9	ug/L	104			78	123	
	Bromoform	20	20.0	ug/L	100			66	130	
	Isopropylbenzene	20	21.0	ug/L	105			72	131	
	1,1,2,2-Tetrachloroethane	20	20.2	ug/L	101			71	121	
	1,3-Dichlorobenzene	20	20.2	ug/L	101			80	119	
	1,4-Dichlorobenzene	20	19.9	ug/L	100			79	118	
	1,2-Dichlorobenzene	20	20.5	ug/L	103			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1559

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045315.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0318WBS01	1,2-Dibromo-3-Chloropropane	20	20.1	ug/L	101			62	128	
	1,2,4-Trichlorobenzene	20	19.7	ug/L	99			69	130	
	1,2,3-Trichlorobenzene	20	19.7	ug/L	99			69	129	