

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559
 Lab File ID: VX045257.D Date Analyzed: 03/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:24
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92968	5.54	162999	6.76	140183	10.05
UPPER LIMIT	185936	6.044	325998	7.257	280366	10.549
LOWER LIMIT	46484	5.044	81499.5	6.257	70091.5	9.549
EPA SAMPLE NO.						
RE115D1-20250310	68915	5.55	136302	6.76	126125	10.05
RE115D2-20250310	67039	5.55	131183	6.76	120316	10.06
RE134D3-20250310	71598	5.55	141729	6.76	129088	10.05
RE134D4-20250310	66430	5.55	130303	6.76	122058	10.06
RE125D3-20250310	69221	5.54	137405	6.76	125442	10.05
RW10-MW01D-20250311	64220	5.54	125900	6.76	116990	10.05
RW10-MW01S-20250311	68205	5.54	136636	6.76	127965	10.05
TT158S1-20250311	70011	5.54	136273	6.76	123646	10.06
TT158I1-20250311	67080	5.54	129858	6.76	122457	10.05
RW10A-MW01S-20250311	68309	5.55	134633	6.76	125955	10.05
RW10A-MW01I-20250311	62583	5.55	123612	6.76	112694	10.06
RW4-RE137-20250311	65951	5.55	131523	6.76	119805	10.05
RE104D1-20250312	71280	5.55	137818	6.76	125719	10.06
TB	67458	5.55	133158	6.76	121095	10.06
RE120D1-20250312	64110	5.54	125932	6.76	116455	10.06
VX0313WBL01	67746	5.54	133586	6.76	123621	10.06
VX0313WBS01	97410	5.55	174469	6.76	153760	10.06
VX0313WBSD01	90235	5.55	159386	6.76	144055	10.05

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559
 Lab File ID: VX045257.D Date Analyzed: 03/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:24
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	92968	5.54	162999	6.76	140183	10.05
UPPER LIMIT	185936	6.044	325998	7.257	280366	10.549
LOWER LIMIT	46484	5.044	81499.5	6.257	70091.5	9.549
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.