

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>			Contract:	<u>AECO15</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1559</u>	SAS No.:	<u>Q1559</u>	SDG NO.:	<u>Q1559</u>
Lab File ID:	<u>VX045257.D</u>			Date Analyzed:	<u>03/13/2025</u>		
Instrument ID:	<u>MSVOA_X</u>			Time Analyzed:	<u>10:24</u>		
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS4 AREA #	RT #				
12 HOUR STD	66961	12.018				
UPPER LIMIT	133922	12.518				
LOWER LIMIT	33480.5	11.518				
EPA SAMPLE NO.						
RE115D1-20250310	51157	12.02				
RE115D2-20250310	49080	12.02				
RE134D3-20250310	55430	12.02				
RE134D4-20250310	52853	12.02				
RE125D3-20250310	54042	12.02				
RW10-MW01D-20250311	47072	12.02				
RW10-MW01S-20250311	52163	12.02				
TT158S1-20250311	50668	12.02				
TT158I1-20250311	51296	12.02				
RW10A-MW01S-20250311	54627	12.02				
RW10A-MW01I-20250311	46139	12.02				
RW4-RE137-20250311	51347	12.02				
RE104D1-20250312	51927	12.02				
TB	50304	12.02				
RE120D1-20250312	49199	12.02				
VX0313WBL01	51498	12.02				
VX0313WBS01	68175	12.02				
VX0313WBSD01	65608	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559
 Lab File ID: VX045257.D Date Analyzed: 03/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:24
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66961	12.018				
UPPER LIMIT	133922	12.518				
LOWER LIMIT	33480.5	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.