

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15  
 Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559  
 Lab File ID: VX045312.D Date Analyzed: 03/18/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 10:02  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                      | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD          | 71356         | 12.018 |  |  |  |  |
| UPPER LIMIT          | 142712        | 12.518 |  |  |  |  |
| LOWER LIMIT          | 35678         | 11.518 |  |  |  |  |
| EPA SAMPLE NO.       |               |        |  |  |  |  |
| RE115D1-20250310DL   | 50596         | 12.02  |  |  |  |  |
| RE134D1-20250310     | 47647         | 12.02  |  |  |  |  |
| RE134D3-20250310DL   | 50569         | 12.02  |  |  |  |  |
| RE134D4-20250310DL   | 52193         | 12.02  |  |  |  |  |
| TT149S1-20250310     | 44828         | 12.02  |  |  |  |  |
| RE125D3-20250310DL   | 39752         | 12.02  |  |  |  |  |
| DUP03-20250310       | 51700         | 12.02  |  |  |  |  |
| RW4-RE137-20250311DL | 57402         | 12.02  |  |  |  |  |
| DUP05-20250312       | 53305         | 12.02  |  |  |  |  |
| RE120D3-20250312     | 58620         | 12.02  |  |  |  |  |
| RE120D2-20250312     | 50819         | 12.02  |  |  |  |  |
| VX0318WBL01          | 46358         | 12.02  |  |  |  |  |
| VX0318WBS01          | 70581         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.