

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559
 Lab File ID: VX045347.D Date Analyzed: 03/20/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:38
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64140	12.018				
UPPER LIMIT	128280	12.518				
LOWER LIMIT	32070	11.518				
EPA SAMPLE NO.						
DUP02-20250310	51206	12.02				
DUP03-20250310DL	45287	12.02				
RE120D3-20250312DL	53223	12.02				
RE120D2-20250312DL	48615	12.02				
VX0320WBL01	47437	12.02				
VX0320WBS01	65210	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.