

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1559 SAS No.: Q1559 SDG NO.: Q1559
Lab File ID: VX045346.D BFB Injection Date: 03/20/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:09
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23
75	30.0 - 60.0% of mass 95	57.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	72.1
175	5.0 - 9.0% of mass 174	5.6 (7.7) 1
176	95.0 - 101.0% of mass 174	68.9 (95.5) 1
177	5.0 - 9.0% of mass 176	4.5 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045347.D	03/20/2025	09:38
VX0320WBL01	VX0320WBL01	VX045349.D	03/20/2025	10:29
VX0320WBS01	VX0320WBS01	VX045350.D	03/20/2025	10:53
DUP03-20250310DL	Q1559-09DL	VX045352.D	03/20/2025	11:44
RE120D2-20250312DL	Q1559-22DL	VX045353.D	03/20/2025	12:07
RE120D3-20250312DL	Q1559-21DL	VX045354.D	03/20/2025	12:30
DUP02-20250310	Q1559-08	VX045356.D	03/20/2025	13:17
VSTDCCC050EC	VSTDCCC050	VX045373.D	03/20/2025	19:51