



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Cradles to Crayons **SDG No.:** Q1561
Contract: CRAD01 **Lab Code:** CHEM **Case No.:** Q1561 **SAS No.:** Q1561
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Copper	2.50	2.0	125	70 - 130	P	03/18/2025	15:58	LB135082
	Lead	0.91	1.0	91	70 - 130	P	03/18/2025	15:58	LB135082