

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Cradles to Crayons</u>	level: <u>low</u>	sdg no.: <u>Q1561</u>
contract: <u>CRAD01</u>	lab code: <u>CHEM</u>	case no.: <u>Q1561</u> sas no.: <u>Q1561</u>
matrix: <u>Water</u>	sample id: <u>Q1561-01</u>	client id: <u>INFANT-ROOM-AMS</u>
Percent Solids for Sample: NA	Spiked ID: Q1561-01MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5150		201		5000	99		P
Lead	ug/L	70 - 130	2240	D	37.7	D	2500	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Cradles to Crayons **level:** low **sdg no.:** Q1561
contract: CRAD01 **lab code:** CHEM **case no.:** Q1561 **sas no.:** Q1561
matrix: Water **sample id:** Q1561-01 **client id:** INFANT-ROOM-AMSD
Percent Solids for Sample: NA **Spiked ID:** Q1561-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	4990		201		5000	96		P
Lead	ug/L	70 - 130	2250	D	37.7	D	2500	88		P