

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Cradles to Crayons

Contract: CRAD01

Lab Code: CHEM Case no.: Q1561

Sas No.: Q1561 SDG No.: Q1561

Instrument ID: P7

Start Date : 03/18/2025

Run Number: LB135082

End Date : 03/18/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                  |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-----------------------------------------------|---|------------------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>165Ho                              | Q | Element<br>209Bi | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1405 | 100                                           |   | 100              |   |         |   |         |   |         |   |
| S2              | S2                 | 1419 | 98                                            |   | 101              |   |         |   |         |   |         |   |
| S3              | S3                 | 1423 | 103                                           |   | 102              |   |         |   |         |   |         |   |
| S4              | S4                 | 1426 | 102                                           |   | 103              |   |         |   |         |   |         |   |
| S5              | S5                 | 1430 | 105                                           |   | 104              |   |         |   |         |   |         |   |
| S6              | S6                 | 1433 | 108                                           |   | 106              |   |         |   |         |   |         |   |
| S7              | S7                 | 1437 | 102                                           |   | 99               |   |         |   |         |   |         |   |
| S8              | S8                 | 1440 | 97                                            |   | 84               |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1528 | 94                                            |   | 93               |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1532 | 95                                            |   | 95               |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1541 | 106                                           |   | 101              |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1546 | 108                                           |   | 103              |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1550 | 105                                           |   | 98               |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1554 | 106                                           |   | 103              |   |         |   |         |   |         |   |
| CRI             | CRI                | 1558 | 105                                           |   | 103              |   |         |   |         |   |         |   |
| PB167140BL      | PB167140BL         | 1604 | 102                                           |   | 101              |   |         |   |         |   |         |   |
| PB167140BS      | PB167140BS         | 1608 | 106                                           |   | 100              |   |         |   |         |   |         |   |
| Q1561-01        | INFANT-ROOM-       | 1611 | 124                                           |   | 130              |   |         |   |         |   |         |   |
| Q1561-01MS      | INFANT-ROOM-       | 1626 | 124                                           |   | 123              |   |         |   |         |   |         |   |
| Q1561-01MSD     | INFANT-ROOM-       | 1629 | 130                                           |   | 131              | * |         |   |         |   |         |   |
| Q1561-01A       | INFANT-ROOM-       | 1633 | 136                                           | * | 137              | * |         |   |         |   |         |   |
| Q1561-02        | WADDLERS-A         | 1639 | 124                                           |   | 126              |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1646 | 111                                           |   | 103              |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1649 | 104                                           |   | 104              |   |         |   |         |   |         |   |
| Q1561-03        | PRE-TODDLER-       | 1653 | 116                                           |   | 115              |   |         |   |         |   |         |   |
| Q1561-04        | LONGE-A            | 1657 | 123                                           |   | 123              |   |         |   |         |   |         |   |
| Q1561-05        | TODDLERS-SIN       | 1701 | 116                                           |   | 118              |   |         |   |         |   |         |   |
| Q1561-06        | STAFF-BATHRO       | 1704 | 112                                           |   | 114              |   |         |   |         |   |         |   |
| Q1561-07        | SICK-BATHRO        | 1708 | 129                                           |   | 133              | * |         |   |         |   |         |   |
| Q1561-08        | DIRECTORS-OF       | 1711 | 121                                           |   | 125              |   |         |   |         |   |         |   |
| Q1561-09        | PRE-K-BATHRO       | 1715 | 120                                           |   | 126              |   |         |   |         |   |         |   |
| Q1561-10        | KINDERGARDEN       | 1719 | 121                                           |   | 124              |   |         |   |         |   |         |   |
| Q1561-11        | 2ND-FLOOR-WA       | 1722 | 116                                           |   | 121              |   |         |   |         |   |         |   |
| Q1561-12        | 1ST-FLOOR-WA       | 1726 | 120                                           |   | 126              |   |         |   |         |   |         |   |

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Cradles to Crayons

Contract: CRAD01

Lab Code: CHEM Case no.: Q1561

Sas No.: Q1561 SDG No.: Q1561

Instrument ID: P7

Start Date : 03/18/2025

Run Number: LB135082

End Date : 03/18/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                  |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-----------------------------------------------|---|------------------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>165Ho                              | Q | Element<br>209Bi | Q | Element | Q | Element | Q | Element | Q |
| CCV03           | CCV03              | 1735 | 110                                           |   | 102              |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1739 | 107                                           |   | 110              |   |         |   |         |   |         |   |
| Q1561-13        | FB                 | 1746 | 108                                           |   | 110              |   |         |   |         |   |         |   |
| Q1561-01        | INFANT-ROOM-       | 1749 | 112                                           |   | 114              |   |         |   |         |   |         |   |
| Q1561-01MS      | INFANT-ROOM-       | 1800 | 117                                           |   | 122              |   |         |   |         |   |         |   |
| Q1561-01MSD     | INFANT-ROOM-       | 1804 | 118                                           |   | 121              |   |         |   |         |   |         |   |
| Q1561-02        | WADDLERS-A         | 1807 | 113                                           |   | 117              |   |         |   |         |   |         |   |
| Q1561-07        | SICK-BATHRO        | 1811 | 111                                           |   | 113              |   |         |   |         |   |         |   |
| Q1561-08        | DIRECTORS-OF       | 1815 | 121                                           |   | 123              |   |         |   |         |   |         |   |
| Q1561-09        | PRE-K-BATHRO       | 1818 | 114                                           |   | 120              |   |         |   |         |   |         |   |
| CCV04           | CCV04              | 1824 | 110                                           |   | 104              |   |         |   |         |   |         |   |
| CCB04           | CCB04              | 1827 | 108                                           |   | 109              |   |         |   |         |   |         |   |
| Q1561-12        | 1ST-FLOOR-WF       | 1832 | 115                                           |   | 117              |   |         |   |         |   |         |   |
| Q1561-01DUP     | INFANT-ROOM-       | 1836 | 134                                           | * | 136              | * |         |   |         |   |         |   |
| Q1561-01L       | INFANT-ROOM-       | 1839 | 114                                           |   | 120              |   |         |   |         |   |         |   |
| Q1561-01A       | INFANT-ROOM-       | 1845 | 124                                           |   | 127              |   |         |   |         |   |         |   |
| Q1561-01DUP     | INFANT-ROOM-       | 1848 | 122                                           |   | 126              |   |         |   |         |   |         |   |
| Q1561-01L       | INFANT-ROOM-       | 1852 | 128                                           |   | 132              | * |         |   |         |   |         |   |
| CCV05           | CCV05              | 1855 | 108                                           |   | 102              |   |         |   |         |   |         |   |
| CCB05           | CCB05              | 1859 | 111                                           |   | 113              |   |         |   |         |   |         |   |

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Cradles to Crayons

Contract: CRAD01

Lab Code: CHEM Case no.: Q1561

Sas No.: Q1561 SDG No.: Q1561

Instrument ID: P7

Start Date : 03/18/2025

Run Number: LB135082

End Date : 03/18/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |         |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-------------------------------------------|---|---------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element                                   | Q | Element | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1405 |                                           |   |         |   |         |   |         |   |         |   |
| S2              | S2                 | 1419 |                                           |   |         |   |         |   |         |   |         |   |
| S3              | S3                 | 1423 |                                           |   |         |   |         |   |         |   |         |   |
| S4              | S4                 | 1426 |                                           |   |         |   |         |   |         |   |         |   |
| S5              | S5                 | 1430 |                                           |   |         |   |         |   |         |   |         |   |
| S6              | S6                 | 1433 |                                           |   |         |   |         |   |         |   |         |   |
| S7              | S7                 | 1437 |                                           |   |         |   |         |   |         |   |         |   |
| S8              | S8                 | 1440 |                                           |   |         |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1528 |                                           |   |         |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1532 |                                           |   |         |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1541 |                                           |   |         |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1546 |                                           |   |         |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1550 |                                           |   |         |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1554 |                                           |   |         |   |         |   |         |   |         |   |
| CRI             | CRI                | 1558 |                                           |   |         |   |         |   |         |   |         |   |
| PB167140BL      | PB167140BL         | 1604 |                                           |   |         |   |         |   |         |   |         |   |
| PB167140BS      | PB167140BS         | 1608 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01        | INFANT-ROOM-       | 1611 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01MS      | INFANT-ROOM-       | 1626 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01MSD     | INFANT-ROOM-       | 1629 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01A       | INFANT-ROOM-       | 1633 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-02        | WADDLERS-A         | 1639 |                                           |   |         |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1646 |                                           |   |         |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1649 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-03        | PRE-TODDLER-       | 1653 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-04        | LONGE-A            | 1657 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-05        | TODDLERS-SIN       | 1701 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-06        | STAFF-BATHRO       | 1704 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-07        | SICK-BATHRO        | 1708 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-08        | DIRECTORS-OF       | 1711 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-09        | PRE-K-BATHRO       | 1715 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-10        | KINDERGARDEN       | 1719 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-11        | 2ND-FLOOR-WA       | 1722 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-12        | 1ST-FLOOR-WA       | 1726 |                                           |   |         |   |         |   |         |   |         |   |

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Cradles to Crayons

Contract: CRAD01

Lab Code: CHEM Case no.: Q1561

Sas No.: Q1561 SDG No.: Q1561

Instrument ID: P7

Start Date : 03/18/2025

Run Number: LB135082

End Date : 03/18/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |         |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-------------------------------------------|---|---------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element                                   | Q | Element | Q | Element | Q | Element | Q | Element | Q |
| CCV03           | CCV03              | 1735 |                                           |   |         |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1739 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-13        | FB                 | 1746 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01        | INFANT-ROOM-       | 1749 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01MS      | INFANT-ROOM-       | 1800 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01MSD     | INFANT-ROOM-       | 1804 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-02        | WADDLERS-A         | 1807 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-07        | SICK-BATHROO       | 1811 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-08        | DIRECTORS-OF       | 1815 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-09        | PRE-K-BATHRO       | 1818 |                                           |   |         |   |         |   |         |   |         |   |
| CCV04           | CCV04              | 1824 |                                           |   |         |   |         |   |         |   |         |   |
| CCB04           | CCB04              | 1827 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-12        | 1ST-FLOOR-WF       | 1832 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01DUP     | INFANT-ROOM-       | 1836 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01L       | INFANT-ROOM-       | 1839 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01A       | INFANT-ROOM-       | 1845 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01DUP     | INFANT-ROOM-       | 1848 |                                           |   |         |   |         |   |         |   |         |   |
| Q1561-01L       | INFANT-ROOM-       | 1852 |                                           |   |         |   |         |   |         |   |         |   |
| CCV05           | CCV05              | 1855 |                                           |   |         |   |         |   |         |   |         |   |
| CCB05           | CCB05              | 1859 |                                           |   |         |   |         |   |         |   |         |   |