

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1567
Project: Pre Treatment Plant 2025	Sample ID: Q1551-02
Client ID: COMPDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	457		454		1	0.49		03/14/2025
TSS	mg/L	+/-5	438		440		1	0.46		03/17/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1567
Project: Pre Treatment Plant 2025	Sample ID: Q1567-01
Client ID: EFFLUENTDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.12		0.12		1	2.49		03/14/2025
Phosphorus, Total	mg/L	+/-20	0.19		0.19		1	1.57		03/14/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1567
Project: Pre Treatment Plant 2025	Sample ID: Q1567-01
Client ID: EFFLUENTMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.55		0.56		1	0.54		03/14/2025
Phosphorus, Total	mg/L	+/-20	0.65		0.65		1	0.46		03/14/2025

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	Q1567
Project:	Pre Treatment Plant 2025	Sample ID:	Q1567-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	32.9		32.9		1	0		03/14/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1567
Project: Pre Treatment Plant 2025	Sample ID: Q1567-05
Client ID: INFLUENTDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	191	OR	191	OR	1	0		03/19/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q1567
Project: Pre Treatment Plant 2025	Sample ID: Q1567-05
Client ID: INFLUENTMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	251	OR	281	OR	1	11		03/19/2025