

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1567

Project: Pre Treatment Plant 2025

Run No.: LB135027

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135027BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	03/14/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1567

Project: Pre Treatment Plant 2025

Run No.: LB135031

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135031BS							
BOD5	mg/L	198	191		96	1	84.6-115.4	03/14/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1567

Project: Pre Treatment Plant 2025

Run No.: LB135037

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135037BS							
Orthophosphate as P	mg/L	0.5	0.51		101	1	90-110	03/14/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1567

Project: Pre Treatment Plant 2025

Run No.: LB135048

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135048BS							
TSS	mg/L	550	533		97	1	90-110	03/17/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1567

Project: Pre Treatment Plant 2025

Run No.: LB135041

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167144BS							
Phosphorus, Total	mg/L	0.50	0.50		99	1	90-110	03/14/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1567

Project: Pre Treatment Plant 2025

Run No.: LB135090

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167189BS							
Ammonia as N	mg/L	1	0.99		99	1	90-110	03/19/2025