

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1569

Project: Fort Meade MD Tipton Airfield Parcel RI - 0111169

RunNo.: LB135044

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/14/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/14/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/14/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/14/2025

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1569

Project: Fort Meade MD Tipton Airfield Parcel RI - 0111169

RunNo.: LB135124

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/21/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/21/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/21/2025